

## CERTIFICATE OF ANALYSIS

<b>Catalog #:</b>	8120	<b>Lot #:</b>	M44314023
<b>Description:</b>	ELISA IgM Assay Diluent 1		
<b>Source:</b>	Product manufactured in the United States using only domestic animal materials processed in U.S. facilities.		
<b>Format:</b>	Liquid Product is 0.2 µm filtered	<b>Exp. Date:</b>	25 Oct 2024
<b>Bioburden:</b>	Acceptable Range: 0 CFU/mL Testing Results: 0 CFU/mL	<b>DOM:</b>	25 Oct 2022
<b>Immunopotency:</b>	Acceptable Range: ± 5% of Standard at a 1:2000 dilution Testing Results: Pass		
<b>Specific Gravity:</b>	Acceptable Range: 1.0483 g/mL – 1.0686 g/mL Testing Results: Pass		
<b>Buffer:</b>	Tris base, Protein Stabilizers, Goat anti-Human IgG, pH 7.4 – 7.6		
<b>Preservative:</b>	0.1% Sodium Azide		
<b>Storage:</b>	Store at 2-8°C		
<b>Inactivation:</b>	None		
<b>Applications:</b>	Blocking buffer. Each laboratory should determine an optimum working titer for use in its particular application. Other applications have not been tested but use in such assays should not necessarily be excluded.		
<b>Safety Note(s):</b>	Refer to the appropriate Safety Data Sheet (SDS) for additional information.		

Quality Signature: 

16 Nov 2022

**FOR RESEARCH OR FURTHER MANUFACTURING USE ONLY**