



Meridian
Life Science, Inc.
Innovative Solutions. Trusted Partner.®

5171 Wilfong Road
Memphis, TN 38134
USA
Telephone: 901-382-8716
Fax: 901-333-8223
Email: info@meridianlifescience.com
www.MeridianLifeScience.com

CERTIFICATE OF ANALYSIS

Important Note: Centrifuge before opening to ensure complete recovery of vial contents.

Catalog #: A32231H **Lot #:** 1L34321

Description: Human CA15-3 LCR Calib. Grade
Human Breast Tumor Antigen CA15-3 Calibrator Grade
Low Cross-Reactivity

Source: Cell Culture Supernatant

Format: Calibrator Grade, Liquid

Purification: Relative Purity Ratio: 228.97 Units/mL/OD280nm
Known Contaminants: 316.1 Units/mL (2.94%) Cobas CA125 Assay
81.59 Units/mL (0.76%) Cobas CA19-9 Assay
Product is 0.2 µm filtered.

Concentration: Specific Protein: 10739 Units/mL Elecsys CA 15-3 Assay
Total Protein: 46.9 OD280nm Units

Buffer: Phosphate Buffered Saline, pH 7.4 ± 0.1

Preservative: 0.1% Sodium Azide

Applications: Suitable for use in Chemiluminescent Immunoassay. Each laboratory should determine an optimum working titer for use in its particular application. Other applications have not been tested but use in such assays should not necessarily be excluded.

Storage: Short-term store at 2–8°C. Long term store at -20° C. Avoid multiple freeze/thaw cycles.

Inactivation: Not Applicable

Safety Note(s): Refer to the appropriate Safety Data Sheet (SDS) for additional information.

25 APR 2022

FOR RESEARCH OR FURTHER MANUFACTURING USE ONLY