



Meridian

Life Science,® Inc.

Innovative Solutions. Trusted Partner.®

5171 Wilfong Road
Memphis, TN 38134
USA

Telephone: 901-382-8716

Fax: 901-333-8223

Email: info@meridianlifescience.com

www.MeridianLifeScience.com

CERTIFICATE OF ANALYSIS

Important Note: Centrifuge before opening to ensure complete recovery of vial contents.

Catalog #: A62934H **Lot #:** 14G19715

Description: Apo AI/B Calibrator 3
Apolipoprotein AI/B Calibrator 3

Source: Defibrinated Human Plasma

Format: Purified, Liquid

Purification: Multi-Step Precipitation
Product is 0.2µm filtered.

Concentration: Total Protein: 3.5 g/dL (Biuret: Cobas-Mira)
Apo A-I: 313.0 mg/dL
Apo B: 307.0 mg/dL (Turbidimetric Assay (Cobas-Mira))

Buffer: 0.1 M Sodium Borate, 0.2 M Sodium Chloride containing proprietary stabilizers.

Preservative: 0.1% Sodium Azide

Applications: Suitable for use in Turbidimetry. Each laboratory should determine an optimum working titer for use in its particular application. Other applications have not been tested but use in such assays should not necessarily be excluded.

Storage: Short term, store at 2–8°C. Long term, store at -20°C. Avoid multiple freeze/thaw cycles.

Inactivation: Not Applicable

Safety Note(s): Refer to the appropriate Safety Data Sheet (SDS) for additional information.

*Sincerely,
S. Orr*

13 February 2020

FOR RESEARCH OR FURTHER MANUFACTURING USE ONLY

003