



## IgY Lab Systems AB

Address: Boksta 505, 744 95 Vittinge, Sweden

Webpage: [www.igylabs.com](http://www.igylabs.com)

E-mail: [info@igylabs.com](mailto:info@igylabs.com)

### Kit Release Certificate of Protein A ELISA

Catalogue number 03-96

Kit lot number:	270611	Expiry Date:	2023-06 <sup>①</sup>
-----------------	--------	--------------	----------------------

Kit Component	Volume	Amount	Lot No.
Microtitre wells coated with chicken anti-Protein A Ab	NA	8x12 wells	270612 <sup>†</sup>
PBS-Tween Tablets	NA	3 pcs	271001
Tween 20 concentrate	3 mL	1 pcs	270613
Protein A reference (0,5 mg/mL)	200 µL	1 pcs	270216
2M Tris Buffer	25 mL	1 pcs	270231
Biotinylated chicken anti-Protein A (100X) Ab	200 µL	1 pcs	264901
Streptavidin-HRP Conjugate (100X)	200 µL	1 pcs	271105
Substrate Solution, TMB/ H <sub>2</sub> O <sub>2</sub> solution	25 mL	1 pcs	270910
Stop Solution	25 mL	1 pcs	270202

<sup>†</sup> Also valid with lot number 270614

Results	Limits	Results
Protein A 10 ng/mL	OD ≥ 1.3	1.82
Blank	OD < 0.15	0.06
QC sample No. 1*	5.76 to 8.64 ng/mL	8.26
QC sample No. 2*	1.44 to 2.16 ng/mL	1.86
QC sample No. 3*	0.43 to 0.65 ng/mL	0.47
IgG-QC sample*	120 to 180 ng/mL	140

\*=At least three of the criteria for the QC samples must be fulfilled.

① Also refer to Quality Statement QS220619

Date: 2022-06-20

Name: Anders Larsson  
Management  
IgY Lab Systems AB

Date: 2022-06-19

Name: Johan Ståhlberg  
Quality Manager  
IgY Lab Systems AB



## QUALITY STATEMENT, QS220619

### REGARDING RE-RELEASE OF KIT BATCH 270611

Kit batch 270611, with an original expiry of 2022-08, was retested on June 19, 2022.

The results indicate excellent performance and signal output, and the retest passed within release specifications. From these results, the expiry date for kit batch 270619 is extended with one (1) year from the test date and the new expiry is thus set to 2023-06.

From today's date the kit is supplied together with a new release certificate (dated 20220619) and this quality statement (QS220619).

The change has been approved by

  
Johan Stålberg, Quality Manager

2022-06-19

Date