

Certificate of Analysis

COA No: CA_BSM-0107

Version: 02

Air Dryable™ Direct RNA/DNA qPCR Blood, 4x

For research or further manufacturing use only

Catalog No:	MDX121	
Lot No:	B337730	
Storage Conditions:	-20°C	
Component Lot No:	425204A	
Expiry date:	May 2027	

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative RT-PCR analysis amplifying a gene from dilutions of mouse RNA under standard conditions. Pass Criteria: Ct profiles must be consistent between test and reference samples with a ± 0.5 Ct variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample. Pass Criteria: Amplification traces must overlay with the negative control.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25 x 10 ⁻⁴ KU DNase I. Pass Criteria: No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: 9.7 x 10 ⁻³ ng/µL RNase <u>Pass Criteria</u> : Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative:

7.21

J. Rahnenführer

Date: 06th May 2025

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822 USA

<u>Germany</u>

Tel: +1 901.382.8716 Tel: +49 (0)3371 60222 00 Fax: +1 901.382.0027 Fax: +49 (0)3371 60222 01