

COA No: CA_BSM-0103-2

Version: 02

Air-Dryable[™] Direct DNA qPCR Saliva, 4x

Catalog No:	MDX130
Lot No:	B334580
Storage Conditions:	-20°C
Component Lot No:	425203A
Expiry date:	April 2027

For research or further manufacturing use only

Quality Control Parameters

Analysis	Specification	Result
Functional	Quantitative real-time PCR analysis amplifying three target genes from a dilution series of mouse cDNA under standard cycling conditions. <u>Pass Criteria</u> : For amplification below Ct 30, Ct profiles must be consistent for test and reference samples with a ± 0.5 Ct variance. For amplification greater than Ct 30, Ct profiles must be consistent for test and reference samples with a ± 1 Ct variance.	Passed
DNA contamination	Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. <u>Pass Criteria</u> : Test sample must amplify in concordance with control sample.	Passed
DNase contamination	DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis. Limit of detection: 6.25 x 10 ⁻⁴ KU DNase I. <u>Pass Criteria</u> : No detectable degradation.	Passed
RNase contamination	Quantitative PCR analysis with high and low RNase standards. Limit of detection: 9.7 x 10 ⁻³ ng/µL RNase <u>Pass Criteria</u> : Test sample must show less RNase activity than the limit of detection.	Passed

QA / QC Representative: They then

X.Chen

Date: 27th March 2025

United Kingdom

Tel: +44 (0)20 8830 5300 Fax: +44 (0)20 8452 2822

<u>USA</u> Tel: +1 901.382.8716 Fax: +1 901.382.0027 Germany

Tel: +49 (0)3371 60222 00 Fax: +49 (0)3371 60222 01

Info@meridianlifescience.com; www.meridianlifescience.com