
	<b>Certificate of Analysis</b>	COA No: CA_BSM-0108
		Version: 02

<b>Air-Dryable™ Direct RNA/DNA qPCR Stool Mix, 4x</b>  For research or further manufacturing use only	Catalog No:	MDX141
	Lot No:	B336070
	Storage Conditions:	-20°C
	Component Lot No:	425204A
	Expiry date:	May 2027

### Quality Control Parameters

Analysis	Specification	Result
Functional	<p>Quantitative real-time PCR analysis amplifying a target gene from a dilution series of mouse RNA under standard cycling conditions.</p> <p><u>Pass Criteria:</u></p> <p>Amplification profile of a 1:10 dilution must be consistent for the test and reference sample within <math>\leq 0.5</math> Cq difference.</p> <p>The end florescence of the 1:10 dilution must be consistent for the test and reference sample within <math>\leq 0.10</math> difference.</p>	Passed
DNA contamination	<p>Quantitative PCR analysis with no template. Presence of <i>E. coli</i> and mouse genomic DNA checked. Test sample must amplify in concordance with control sample.</p> <p><u>Pass Criteria:</u></p> <p>Amplification traces must overlay with the negative control.</p>	Passed
DNase contamination	<p>DNase contamination is measured as DNA substrate degradation against a DNase I dilution series by agarose gel electrophoresis.</p> <p>Limit of detection: <math>6.25 \times 10^{-4}</math> KU DNase I.</p> <p><u>Pass Criteria:</u></p> <p>No detectable degradation.</p>	Passed
RNase contamination	<p>Quantitative PCR analysis with high and low RNase standards.</p> <p>Limit of detection: <math>9.7 \times 10^{-3}</math> ng/<math>\mu</math>L RNase</p> <p><u>Pass Criteria:</u></p> <p>Test sample must show less RNase activity than the limit of detection.</p>	Passed

QA / QC Representative:



J. Rahnenführer

Date: 15<sup>th</sup> April 2025

**United Kingdom**

Tel: +44 (0)20 8830 5300  
Fax: +44 (0)20 8452 2822

**USA**

Tel: +1 901.382.8716  
Fax: +1 901.382.0027

**Germany**

Tel: +49 (0)3371 60222 00  
Fax: +49 (0)3371 60222 01