

Test Release Document

Released:
Michele Amasio

Version: 01

Lyo-Ready[™] Genotyping Direct qPCR Urine

Cat. No. MD158

Storage Conditions: -20 °C

Lot number: RD123F

Test Parameters

Analysis	Specification	Result
Functionality	Quantitative PCR analysis amplifying a multiplex reaction of 2 genes from a dilution series of template under standard conditions.	Passed
	Relative activity is measured against a reference by a functional qPCR assay.	
	Test Criteria: - Test and reference amplification profiles must align; ΔCt (Test-Reference) < 1	

Authorised by Thomas Fouqueau

Team Leader R&D